

# CE Declaration of Conformity

We herewith confirm the following designated products to comply with the requirements set out in the Council Directive on the approximation of the laws of the Member States relating to Electromagnetic Compatibility Directive (93/42/EEC) with applicable standards listed below.

Product : Medical grade box PC  
Trade name : WINCOMM  
Model Number : WPC-766X (X=0~9,A~Z or Blank)  
Applicable Harmonized : EN 55011: 2009/A1: 2010 Class B (Group 1)  
Standards under Directive 93/42/EEC : EN 60601-1-2: 2015  
IEC 60601-1-2: 2014  
EN 61000-3-2: 2014  
EN 61000-3-3: 2013

Company Name : \_\_\_\_\_  
Company Address : \_\_\_\_\_  
Telephone : \_\_\_\_\_ Facsimile : \_\_\_\_\_

Person in responsible for marking this declaration:

_____	_____
Name (Full Name)	Title/ Department
_____	_____
Date	Legal Signature



Date : Jan. 15, 2018

---

## CE Statement of Conformity

---

This statement is to certify that the designated product below.

Product : Medical grade box PC  
Trade name : WINCOMM  
Model Number : WPC-766X (X=0~9,A~Z or Blank)  
Company Name : WINCOMM CORPORATION  
Applicable Standards : EN 55011: 2009/A1: 2010 Class B (Group 1)  
EN 60601-1-2: 2015  
IEC 60601-1-2: 2014  
EN 61000-3-2: 2014  
EN 61000-3-3: 2013

One sample of the designated product has been tested and evaluated in our laboratory to find in compliance with the applicable standards above. The issued test report(s) show(s) it in detail.

**Report Number : 17C0023R-ITCEP13V00**

TEST LABORATORY

A handwritten signature in black ink, appearing to read "Vincent Lin", written over a horizontal line.

Vincent Lin / Director

The verification is based on a single evaluation of one sample of above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. Logo.

# CE Test Report

Product Name : Medical grade box PC  
Model No. : WPC-766X (X=0~9,A~Z or Blank)

Applicant : WINCOMM CORPORATION  
Address : 3F., No.14, Zhanye 2nd Rd., East Dist., Hsinchu City 300,  
Taiwan (R.O.C.)

Date of Receipt : 2017/12/04  
Issued Date : 2018/01/15  
Report No. : 17C0023R-ITCEP13V00  
Report Version : V1.0



The test results relate only to the samples tested.  
The test results shown in the test report are traceable to the national/international standard through the calibration of the equipment and evaluated measurement uncertainty herein.  
This report must not be used to claim product endorsement by TAF or any agency of the government.  
The test report shall not be reproduced without the written approval of DEKRA Testing and Certification Co., Ltd.

# Test Report

Issued Date : 2018/01/15  
Report No. : 17C0023R-ITCEP13V00



Product Name : Medical grade box PC  
 Applicant : WINCOMM CORPORATION  
 Address : 3F., No.14, Zhanye 2nd Rd., East Dist., Hsinchu City 300,  
 Taiwan (R.O.C.)  
 Manufacturer : WINCOMM CORPORATION  
 Model No. : WPC-766X (X=0~9,A~Z or Blank)  
 EUT Rated Voltage : AC 100-240V, 47-63Hz  
 EUT Test Voltage : AC 230V/50Hz  
 Trade Name : WINCOMM  
 Applicable Standard : EN 55011: 2009/A1: 2010 Class B (Group 1)  
 EN 60601-1-2: 2015  
 IEC 60601-1-2: 2014  
 EN 61000-3-2: 2014  
 EN 61000-3-3: 2013

Test Result : Complied  
 Performed Location : DEKRA Testing and Certification Co., Ltd.  
 Linkou Laboratory  
 No. 5-22, Ruishukeng  
 Linkou District, New Taipei City, 24451, Taiwan  
 TEL:+866-2-8601-3788 / FAX:+886-2-8601-3789

Documented By : Leven Huang  
 (Senior Adm. Specialist / Leven Huang )

Reviewed By : Eddie Tseng  
 ( Engineer / Eddie Tseng )

Approved By : Vincent Lin  
 ( Director / Vincent Lin)

## Laboratory Information

We, **DEKRA Testing and Certification Co., Ltd.**, are an independent EMC and safety consultancy that was established the whole facility in our laboratories. The test facility has been accredited/accepted (audited or listed) by the following related bodies in compliance with ISO 17025, EN 45001 and specified testing scopes:

<b>Taiwan R.O.C.</b>	<b>:</b>	<b>BSMI, NCC, TAF</b>
<b>Norway</b>	<b>:</b>	<b>DNV</b>
<b>USA</b>	<b>:</b>	<b>FCC</b>
<b>Japan</b>	<b>:</b>	<b>VCCI</b>

The related certificate for our laboratories about the test site and management system can be downloaded from DEKRA Testing and Certification Co., Ltd. Web Site :

<http://www.dekra.com.tw/english/about/certificates.aspx?bval=5>

The address and introduction of DEKRA Testing and Certification Co., Ltd. laboratories can be founded in our Web site : [http://www.dekra.com.tw/index\\_en.aspx](http://www.dekra.com.tw/index_en.aspx)

If you have any comments, Please don't hesitate to contact us. Our contact information is as below:

### Hsin Chu Laboratory:

No.75-2, 3rd Lin, WangYe Keng, Yonghxing Tsuen, Qionglin Shiang, Hsinchu County 307, Taiwan (R.O.C.)

No.372, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County 310, Taiwan

No.372-2, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County 310, Taiwan

TEL:+886-3-592-8858 / FAX:+886-3-592-8859

TEL:+886-3-582-8001 / FAX:+886-3-5828-958 E-Mail : [info.tw@dekra.com](mailto:info.tw@dekra.com)

### Lin Kou Laboratory

No. 5-22, Ruishukeng, Linkou Dist., New Taipei City 24451, Taiwan (R.O.C.)

TEL : +886-2-8601-3788 / FAX : +886-2-8601-3789 E-Mail : [info.tw@dekra.com](mailto:info.tw@dekra.com)

### Suzhou (China) Testing Laboratory :

No.99 Hongye Rd., Suzhou Industrial Park, Suzhou,215006, Jiangsu,China

TEL : +86-512-6251-5088 / FAX : +86-512-6251-5098 E-Mail : [info.tw@dekra.com](mailto:info.tw@dekra.com)