

CE Declaration of Conformity

We herewith confirm the following designated products to comply with the requirements set out in the Council Directive on the approximation of the laws of the Member States relating to Electromagnetic Compatibility Directive (93/42/EEC) with applicable standards listed below.

Product : Medical Panel PC
Trade name : WINCOMM
Model Number : WMP-24C, WMP-24F
Applicable Harmonized : EN 55011: 2009/A1: 2010 Class B (Group 1)
Standards under Directive EN 60601-1-2: 2015
93/42/EEC IEC 60601-1-2: 2014
EN 61000-3-2: 2014
EN 61000-3-3: 2013

Company Name : _____

Company Address : _____

Telephone : _____ Facsimile : _____

Person in responsible for marking this declaration:

Name (Full Name)

Title/ Department

Date

Legal Signature



Date : July. 13, 2017

CE Statement of Conformity

This statement is to certify that the designated product below.

Product : Medical Panel PC
Trade name : WINCOMM
Model Number : WMP-24C, WMP-24F
Company Name : WINCOMM CORPORATION
Applicable Standards : EN 55011: 2009/A1: 2010 Class B (Group 1)
EN 60601-1-2: 2015
IEC 60601-1-2: 2014
EN 61000-3-2: 2014
EN 61000-3-3: 2013

One sample of the designated product has been tested and evaluated in our laboratory to find in compliance with the applicable standards above. The issued test report(s) show(s) it in detail.

Report Number : 1760190R-ITCEP28V00

TEST LABORATORY

A handwritten signature in black ink, appearing to read 'Vincent Lin', written over a horizontal line.

Vincent Lin / Director

The verification is based on a single evaluation of one sample of above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. Logo.

CE Test Report

Product Name : Medical Panel PC
Model No. : WMP-24C, WMP-24F

Applicant : WINCOMM CORPORATION
Address : 3F., No.14, Zhanye 2nd Rd., East Dist., Hsinchu City 300,
Taiwan (R.O.C.)

Date of Receipt : 2017/06/06
Issued Date : 2017/07/13
Report No. : 1760190R-ITCEP28V00
Report Version : V1.0



The test results relate only to the samples tested.
The test results shown in the test report are traceable to the national/international standard through the calibration of the equipment and evaluated measurement uncertainty herein.
This report must not be used to claim product endorsement by TAF or any agency of the government.
The test report shall not be reproduced without the written approval of DEKRA Testing and Certification Co., Ltd.

Test Report

Issued Date : 2017/07/13
Report No. : 1760190R-ITCEP28V00



Product Name : Medical Panel PC
Applicant : WINCOMM CORPORATION
Address : 3F., No.14, Zhanye 2nd Rd., East Dist., Hsinchu City 300,
Taiwan (R.O.C.)
Manufacturer : WINCOMM CORPORATION
Model No. : WMP-24C, WMP-24F
EUT Rated Voltage : AC 100-240V, 47-63Hz
EUT Test Voltage : AC 230V/50Hz
Trade Name : WINCOMM
Applicable Standard : EN 55011: 2009/A1: 2010 Class B (Group 1)
EN 60601-1-2: 2015
IEC 60601-1-2: 2014
EN 61000-3-2: 2014
EN 61000-3-3: 2013
Test Result : Complied
Performed Location : DEKRA Testing and Certification Co., Ltd.
Linkou Laboratory
No. 5-22, Ruishukeng
Linkou District, New Taipei City, 24451, Taiwan
TEL:+866-2-8601-3788 / FAX:+886-2-8601-3789

Documented By : Leven Huang
(Senior Adm. Specialist / Leven Huang)

Reviewed By : Sam Lin
(Engineer / Sam Lin)

Approved By : Vincent Lin
(Director / Vincent Lin)

Laboratory Information

We, **DEKRA Testing and Certification Co., Ltd.**, are an independent EMC and safety consultancy that was established the whole facility in our laboratories. The test facility has been accredited/accepted (audited or listed) by the following related bodies in compliance with ISO 17025, EN 45001 and specified testing scopes:

Taiwan R.O.C.	:	BSMI, NCC, TAF
Norway	:	DNV
USA	:	FCC
Japan	:	VCCI

The related certificate for our laboratories about the test site and management system can be downloaded from DEKRA Testing and Certification Co., Ltd. Web Site :

<http://www.dekra.com.tw/english/about/certificates.aspx?bval=5>

The address and introduction of DEKRA Testing and Certification Co., Ltd. laboratories can be founded in our Web site : http://www.dekra.com.tw/index_en.aspx

If you have any comments, Please don't hesitate to contact us. Our contact information is as below:

Hsin Chu Laboratory:

No.75-2, 3rd Lin, WangYe Keng, Yonghxing Tsuen, Qionglin Shiang, Hsinchu County 307, Taiwan (R.O.C.)

No.372, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County 310, Taiwan

No.372-2, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County 310, Taiwan

TEL:+886-3-592-8858 / FAX:+886-3-592-8859

TEL:+886-3-582-8001 / FAX:+886-3-5828-958 E-Mail : info.tw@dekra.com

Lin Kou Laboratory

No. 5-22, Ruishukeng, Linkou Dist., New Taipei City 24451, Taiwan (R.O.C.)

TEL : +886-2-8601-3788 / FAX : +886-2-8601-3789 E-Mail : info.tw@dekra.com

Suzhou (China) Testing Laboratory :

No.99 Hongye Rd., Suzhou Industrial Park, Suzhou,215006, Jiangsu,China

TEL : +86-512-6251-5088 / FAX : +86-512-6251-5098 E-Mail : info.tw@dekra.com