

CE Declaration of Conformity

We herewith confirm the following designated products to comply with the requirements set out in the Council Directive on the approximation of the laws of the Member States relating to Electromagnetic Compatibility Directive (93/42/EEC) with applicable standards listed below.

Product : Medical Panel PC
Trade name : WINCOMM
Model Number : WMP-17F
Applicable Harmonized : EN 55011: 2009+A1: 2010 Class B (Group 1)
Standards under Directive EN 60601-1-2: 2015
93/42/EEC IEC 60601-1-2: 2014
EN 61000-3-2: 2014
EN 61000-3-3: 2013

Company Name : _____
Company Address : _____
Telephone : _____ Facsimile : _____

Person in responsible for marking this declaration:

_____ Name (Full Name)	_____ Title/ Department
_____ Date	_____ Legal Signature

CE Statement of Conformity

This statement is to certify that the designated product below.

Product : Medical Panel PC
Trade name : WINCOMM
Model Number : WMP-17F
Company Name : WINCOMM CORPORATION
Applicable Standards : EN 55011: 2009+A1: 2010 Class B (Group 1)
EN 60601-1-2: 2015
IEC 60601-1-2: 2014
EN 61000-3-2: 2014
EN 61000-3-3: 2013

One sample of the designated product has been tested and evaluated in our laboratory to find in compliance with the applicable standards above. The issued test report(s) show(s) it in detail.

Report Number : 1690020R-ITCEP13V00

TEST LABORATORY



Vincent Lin / Director

The verification is based on a single evaluation of one sample of above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. Logo.

CE Test Report

Product Name : Medical Panel PC

Model No. : WMP-17F

Applicant : WINCOMM CORPORATION

Address : 3F, No.14, Prosperity Road II, Science-Based
Industrial Park, Hsinchu Taiwan 30078, R.O.C

Date of Receipt : 2016/08/26

Issued Date : 2016/09/08

Report No. : 1690020R-ITCEP13V00

Report Version : V1.0



The test results relate only to the samples tested.

The test results shown in the test report are traceable to the national/international standard through the calibration of the equipment and evaluated measurement uncertainty herein.

This report must not be used to claim product endorsement by TAF or any agency of the government.

The test report shall not be reproduced without the written approval of QuieTek Corporation.

Test Report

Issued Date : 2016/09/08
Report No. : 1690020R-ITCEP13V00



Product Name : Medical Panel PC
Applicant : WINCOMM CORPORATION
Address : 3F, No.14, Prosperity Road II, Science-Based Industrial
Park, Hsinchu Taiwan 30078, R.O.C
Manufacturer : WINCOMM CORPORATION
Model No. : WMP-17F
EUT Rated Voltage : AC 100-240V, 50/60Hz
EUT Test Voltage : AC 230V/50Hz
Trade Name : WINCOMM
Applicable Standard : EN 55011: 2009+A1: 2010 Class B (Group 1)
EN 60601-1-2: 2015
IEC 60601-1-2: 2014
EN 61000-3-2: 2014
EN 61000-3-3: 2013
Test Result : Complied
Performed Location : Quietek Corporation (Linkou Laboratory)
No. 5-22, Ruishukeng, Linkou Dist., New Taipei City 24451,
Taiwan. R.O.C.
TEL:+866-2-8601-3788 / FAX:+886-2-8601-3789

Documented By : April Chen
(Adm. Specialist / April Chen)

Tested By : Fran Chen
(Engineer / Fran Chen)

Approved By : [Signature]
(Director / Vincent Lin)

Laboratory Information

We, **Quietek Corporation**, are an independent EMC and safety consultancy that was established the whole facility in our laboratories. The test facility has been accredited/accepted (audited or listed) by the following related bodies in compliance with ISO 17025, EN 45001 and specified testing scopes:

Taiwan R.O.C.	:	BSMI, NCC, TAF
Norway	:	DNV
USA	:	FCC
Japan	:	VCCI

The related certificate for our laboratories about the test site and management system can be downloaded from Quietek Corporation's Web Site : <http://www.quietek.com/chinese/about/certificates.aspx?bval=5>

The address and introduction of Quietek Corporation's laboratories can be founded in our Web site : <http://www.quietek.com/>

If you have any comments, Please don't hesitate to contact us. Our contact information is as below:

HsinChu Testing Laboratory :

No.75-2, 3rd Lin, Wangye Keng, Yonghxing Tsuen, Qionglin Shiang, Hsinchu County 307, Taiwan, R.O.C.

TEL:+886-3-592-8858 / FAX:+886-3-592-8859 E-Mail : service@quietek.com

LinKou Testing Laboratory :

No.5-22, Ruishukeng, Linkou Dist., New Taipei City 24451, Taiwan, R.O.C.

TEL : 886-2-8601-3788 / FAX : 886-2-8601-3789 E-Mail : service@quietek.com

Suzhou (China) Testing Laboratory :

No. 99 Hongye Rd., Suzhou Industrial Park Loufeng Hi-Tech Development Zone., Suzhou,China.

TEL : +86-512-6251-5088 / FAX : +86-512-6251-5098 E-Mail : service@quietek.com